


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-B1-W-P

S.N.: 3501

Date: JAN 23 87



- a) Output signal amplitude:  
0 TO +20 VOLTS
- b) Pulse width:  
0.1 TO 5.0  $\mu$ SEC
- c) Rise time:  
 $\leq 5$  NSEC
- d) Fall time:  
 $\leq 5$  NSEC
- e) PRF:  
0 TO 50 KHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+24  $\pm$  1.0 VOLT  
300 mA (ON VOLT)

$R_L = 50 \sim$   
50 KHZ