


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MP-10-2-75-10*

S.N.: *3511*

Date: *JAN 12 '87*



*40 db ATTEN
10 VOLTS/DIV
50 KHz*



*40 db ATTEN
10 VOLTS/DIV
50 KHz*

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
5 TO 50 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 50 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz