


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-3-PS-P*

S.N.: *3533*

Date: *MAR 4 '87*



*40 db ATTEN*

*10 VOLTS/DIV*

*100 MHz*

- a) Output signal amplitude:  
*0 TO +20 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 300 PSEC*
- d) Fall time:  
*≤ 300 PSEC*
- e) PRF:  
*0 TO 10 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 - 240 V*  
*50 - 60 HZ*

*[Signature]*