


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-A1-PS-05-10A-1*

S.N.: *3534*

Date: *FEB 26 87*



*R_L = 10K.
50 kHz.*

- a) Output signal amplitude:
0 TO 1100 VOLTS TO
- b) Pulse width: *> 10K
50 NSEC TO 1.0 USEC.*
- c) Rise time:
≤ 5 NSEC
- d) Fall time:
≤ 5 NSEC
- e) PRF:
0 TO 200 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60Hz.*