


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-AP1-PS-05-14A-N*

S.N.: *3536*

Date: *FEB 26 87*



*R<sub>L</sub> = 10k.  
50 kHz*

- a) Output signal amplitude:  
*0 TO -100 VOLTS.*
- b) Pulse width: *TO > 10K*
- c) Rise time: *50 NSEC TO 18 USEC*  
*≤ 5 NSEC*
- d) Fall time: *≤ 5 NSEC*
- e) PRF: *0 TO 200 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*120/240 V  
50-60 Hz*

