


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-107B-PS-P*
S.N.: *3544*
Date: *APRIL 30 87*



- a) Output signal amplitude:
0 TO +2.0 AMPS
- b) Pulse width: *(0 TO 30 μ)*
20 TO 200 NS
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

R_L = 30 Ω
 ≈ 0.6 AMP/DIV
20 KHz

[Signature]