


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PM-2-EN-7*

S.N.: *3586*

Date: *JAN 21 87*



- a) Output signal amplitude:
0 TO 15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz

f) Jitter, stability:

OK

- g) Prime power:
*+24 ± 1.0 VOLTS,
500 mA MAX.*

*40 db ATTEN
10 VOLTS/DIV
7.0 MHz*

[Signature]