


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-U-P*

S.N.: *3603*

Date: *DEC 23 86*



40 db ATTEN
5 VOLTS/DIV
50 KHz

- a) Output signal amplitude:
≥ +5 VOLTS
- b) Pulse width:
2 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 1 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT
250 mA MAX