

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVK-V P

S.N.:

3605

Date:

DEC 23 86



40 dB ATTEN
5 VOLTS/DIV
50 kHz

a) Output signal amplitude:

$\approx \pm 5$ VOLTS

b) Pulse width:

1 TO 50 NSEC.

c) Rise time:

≤ 100 PSEC.

d) Fall time:

≤ 1 NSEC

e) PRF:

0 TO 50 KHz.

f) Jitter, stability:

OK

g) Prime power:

$\pm 15 \pm 1.0$ VOLT.

250 mA.

(MAX)

[Signature]