


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV14-MART-P

S.N.: 3606

Date: MARCH 3 87



50 db ATTEN
500 kHz

a) Output signal amplitude:

$\approx +30$ VOLTS

b) Pulse width:

≤ 1.0 nSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0$ V

150 mA (MAX)

[Signature]