


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AW-1-P*
S.N.: *3613*
Date: *DEC 8 '86*



- a) Output signal amplitude: *20 TO 100 AMPS*
- b) Pulse width: *25 USEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1KHZ*
- f) Jitter, stability: *OK*
- g) Prime power:

100 AMPS/DIV.
10 USEC/DIV
1 KHZ

+15 ± 1.0 VOLT
250 mA
(MAX)
SB