


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5-C-P*

S.N.: *3616*

Date: *FEB 13 85*



- a) Output signal amplitude:  
*0 TO +25 AMPS*
- b) Pulse width:  
*5 TO 100 NSEC.*
- c) Rise time:  
*≤ 3 NSEC*
- d) Fall time:  
*≤ 4 NSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*

*100. VOLTS/DIV*  
*10 NSEC/DIV*  
*R<sub>L</sub> = 3 Ω*  
*2.0 KHz*

g) Prime power:  
*120/240 V*  
*50-60 Hz*  
*AA*