


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-F*

S.N.: *3620*

Date: *MAR 3 87*



50 dB ATTEN

10.0 KHz

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
2 TO 100 nSEC
- c) Rise time:
≤ 350 pSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
MIN PW: 0 TO 1.0 MHz
MAX PW: 0 TO 20 KHz
- f) Jitter, stability:
OK.
- g) Prime power:
+15 ± 1.0 VOLT
150 mA (MAX)