


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MH-50PT-P*  
S.N.: *3634*  
Date: *JAN 19 87*



- a) Output signal amplitude:  
*> +30 VOLT*
- b) Pulse width:  
*≤ 1.0 NS*
- c) Rise time:  
*N/A*
- d) Fall time:  
*N/A*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 dB*  
*32 VOLT/DIV*  
*1.0 MHz*

*+15 ± 1.0 VOLTS*  
*150 mA (MAX)*  
*[Signature]*