


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *3635*

Date: *JAN 19 87*



- a) Output signal amplitude:
> +30 VOLTS.
- b) Pulse width:
≤ 1.0 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

50 dB
32 VOLTS DIV
1.0 MHz

+15 ± 1.0 VOLTS
150 mA (MAX)
[Signature]