


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-1-PW-C-7*

S.N.: *3649*

Date: *FEB 13 87*



*$R_L = 50 \Omega$
1 KHz*


- a) Output signal amplitude: *0.7V ± 200.40V*
- b) Pulse width: *0.17V 5.0 μs*
- c) Rise time: *≤ 10 ns*
- d) Fall time: *≤ 10 ns*
- e) PRF: *0.7V 1 KHz*
- f) Jitter, stability:
- g) Prime power: *OK*

*120/240V
50-60Hz*

[Signature]

PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-1-pw-C-N*
S.N.: *3649 (MOP)*
Date: *MAY 28 87*



- a) Output signal amplitude:
0 TO -200 VOLTS
- b) Pulse width:
0.1 TO 5.0 μS
- c) Rise time:
≤ 10 nS
- d) Fall time:
≤ 10 nS
- e) PRF:
0 TO 1.0 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

R_L = 50 Ω
1.0 kHz