


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ANMN-1-C-PN*

S.N.: *3656*

Date: *APRIL 16 '87*



Ⓐ 40 dB ATTEN
5 VOLTS/DIV
15 MHz



Ⓑ AS Ⓐ BUT
≈ 50 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.3 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 HZ