


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *PVRL-1TT5D*

S.N.: *3669*

Date: *MAR 23 87*



*A OUT*



*B OUT.*

- a) Output signal amplitude:
  - A) -200 V TO > 10K*
  - B) -200 V TO > 10K*
- b) Pulse width:
  - A) 20 TO 100 NSEC*
  - B) 100 NSEC TO 5 MSEC*
- c) Rise time:
  - A) < 5 NSEC*
  - B) < 10 NSEC*
- d) Fall time:
  - A) < 5 NSEC*
  - B) < 10 NSEC*
- e) PRF:
  - 0 TO 50 Hz.*
- f) Jitter, stability:
  - OK.*
- g) Prime power:
  - 120/240 V*
  - 50-60 Hz.*

*[Signature]*