


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVK-AN-PS-P*

S.N.: *3686*

Date: *APRIL 29 87*



*ⓐ 40 db ATTEN
10 VOLTS/DIV
20 kHz*



*ⓑ AS ⓐ BUT
1 nSEC/DIV
(RISE TIME)*

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
2 TO 50 nSEC
- c) Rise time:
≤ 100 nSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
ok
- g) Prime power:

*120/240V
50-60 Hz*