


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5-CP*

S.N.: *3696*

Date: *MAY 4 87*



*~ 33 AMP/Div  
10 nSEC/Div  
2.0 kHz*

- a) Output signal amplitude:  
*0 TO +25 AMP'S*
- b) Pulse width:  
*5 TO 100 nSEC*
- c) Rise time:  
*~ 3 nSEC*
- d) Fall time:  
*~ 4 nSEC*
- e) PRF:  
*0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V  
50-60 Hz*

*[Signature]*