


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5-C-P*  
S.N.: *3697*  
Date: *MAY 4 87*



- a) Output signal amplitude:  
*0 TO 120 mV*
- b) Pulse width:  
*5 TO 100 nSEC*
- c) Rise time:  
*≤ 3 nSEC*
- d) Fall time:  
*≤ 4 nSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*

*33 AMP/DIV*  
*10 nSEC/DIV*  
*2.0 KHz*

g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*