


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-C-05-P*

S.N.: *3701*

Date: *MAY 12 87*



- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≈ 4 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN
100 VOLTS/DIV
50 KHz

170 / 240V
50-60 Hz