


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVP-AV-15-C-P-PN-M*

S.N.: *3710*

Date: *MAY 11 87*



*40 dB ATTEN  
5 VOLTS/DIV  
1.0 MHz*



*40 dB ATTEN.  
RISE TIME*

- a) Output signal amplitude:  
*0 TO 15 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NSER.*
- c) Rise time:  
*≤ 60 PSEZ*
- d) Fall time:  
*≤ 130 PSEZ.*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V  
50-60 Hz.*

*[Signature]*