


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMMN-3A-C-PN*

S.N.: *3714*

Date: *MAY 13 '87*



*Ⓐ 40 db ATTEN
5 VOLTS/DIV
≈ 10 MHz*



*Ⓑ AS Ⓐ BUT
≈ 50 MHz*

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
5 TO 40 NS
- c) Rise time: *6 / NS*
(30% MAX DUTY CYCLE)
- d) Fall time:
6 / NS
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240
50 - 60 Hz

[Signature]