



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2A-C-P-10SC1
S.N.: 3723
Date: APRIL 10 '85



Ⓐ ≈ 7 AMPS/DIV
20 KHz



Ⓑ ≈ 7 AMPS/DIV
20 KHz
FALL TIME
 ≤ 1.0 NSEC (20-80%)

- a) Output signal amplitude:
0 TO +5 AMPS
- b) Pulse width:
2 TO 50 NSEC
- c) Rise time:
 ≤ 1 NSEC
- d) Fall time:
 ≤ 1 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

[Signature]