


PULSE GENERATOR
PERFORMANCE CHECK


Model: AMR-133-PS-PN-W

S.N.: 3733

Date: JUNE 30¹ 1970



Ⓐ P out.
 $P_o = 50\text{m}$
5 KHz



Ⓑ N out
 $P_o = 50\text{m}$
5 KHz

- a) Output signal amplitude:
0 TO ±100 VOLTS
- b) Pulse width:
0.1 TO 50 μSEC
- c) Rise time:
≤ 5 nSEC
- d) Fall time:
≤ 5 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

SA