


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MM-3-P*
S.N.: *3739*
Date: *JUNE 4 87*



40 dB atten
10 VOLT/DIV
1.0 MHz

- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+28 ± 2 VOLTS