


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AU-1A3-C-P*

S.N.: *3753*

Date: *JUNE 5 87*



*50 db ATTEN*  
*2.32 VOLTS/DIV*  
*100 KHz*

- a) Output signal amplitude:  
*0 TO +40 VOLTS*
- b) Pulse width:  
*0.4 TO 2.0 NS*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 250 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*