


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-101-C-PN

S.N.: 3756

Date: JUNE 2 87



60 db ATTEN
100 VOLTS/DIV
100 KHz

- a) Output signal amplitude:
0 TO \pm 100 VOLTS
- b) Pulse width:
1.0 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz