


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMAN-1-C-NOT-M*
S.N.: *3703*
Date: *JULY 20 '87*



40 db ATTEN
0.5 VOLT/DIV
12 MHz

- a) Output signal amplitude:
0 TO - 5 VOLTS
- b) Pulse width:
0.2 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC.
- e) PRF:
0 TO 50 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.