

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM14-2 PS-PN-M*

S.N.: *3784*

Date: *JUN 9 87*



*50 dB ATTEN*

*.2 32 VOLTS/DIV*

*25 MHz*

- a) Output signal amplitude:  
*0 TO ± 30 VOLTS*
- b) Pulse width:  
*≤ 400 PSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*