


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-1-P*
S.N.: *3785*
Date: *July 8 87*



R_L = 50 Ω .
1.0 kHz.

- a) Output signal amplitude:
0 TO +200 V
- b) Pulse width:
0.1 TO 5.0 μ SEC
- c) Rise time:
 \leq 10 NSER
- d) Fall time:
 \leq 10 NSER
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.0 VOLTS
500 mA (MAX)