


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AM-1-C-PN*
S.N.: *3791*
Date: *JULY 30 87*



40 dB ATTEN
50 VOLTS/DIV
12 MHz

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz
[Signature]