


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U.C-P*
S.N.: *3797*
Date: *MAY 8 87*



50 db ATTEN
≈ 32 VOLTS/DIV
20 KHZ

- a) Output signal amplitude:
0 TO + 50 VOLTS
- b) Pulse width:
2 TO 100 NSEC
- c) Rise time:
≤ 350 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ