


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PNH-MPT-P*

S.N.: *3B46*

Date: *Jul 5 '87*



- a) Output signal amplitude:  
*≥ 30 VOLTS*
- b) Pulse width:  
*≤ 1.0 NSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLTS*

*150 mA (MAX)*

*[Signature]*

*50 dB ATTEN*

*32 VOLTS/DIV*

*500 KHz*