


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PM1-MPT-P*

S.N.: *3840*

Date: *Aug 5 '87*



- a) Output signal amplitude: *30 VOLTS.*
- b) Pulse width: *≤ 1.0 nSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLTS*

50 db ATTEN
32 VOLTS/DIV.
5.00 kHz

150mA (MAX.)
[Signature]