


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVH-M2-P*

S.N.: *3851*

Date: *Nov 26 87*



Ⓐ 50 db ATTEN
≈ 32 VOLTS/DIV
500 KHz



Ⓑ AS Ⓐ BUT
100 PSEC/DIV
(JITTER CHECK)

- a) Output signal amplitude:
≥ +60 VOLTS
- b) Pulse width:
≤ 2 nSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+135 ± 5 VOLTS
5 mA (MAX)