


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APR-AY-PW-EA 7*  
S.N.: *3852*  
Date: *Jul 5 '87*



- a) Output signal amplitude:  
*0 TO +200V.*
- b) Pulse width:  
*0.05 TO 100 NSER.*
- c) Rise time:  
*≤ 10 NSER*
- d) Fall time:  
*≤ 10 NSER*
- e) PRF:  
*MIN PW: 0 TO 100 KHZ*  
*MAX PW: 0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*424 ± 1.0 VOUT3*  
*500 mA (MAX)*

*R<sub>L</sub> = 50 Ω*  
*5 KHZ*