


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVL-2-C-CA-C-M-05-PN
S.N.: 3853
Date: AUG 13 1987



70 db ATTEN
≈ 320 VOUT/DIV
2.0 kHz

- a) Output signal amplitude:
0 to ±350 VOLTS
- b) Pulse width:
5 to 100 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 kHz