


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PK-A-PN*  
S.N.: *3867*  
Date: *APR 12 87*



*40 dB ATTEN*

*10 VOLTS/DIV*

*1.0 MHz*

- a) Output signal amplitude:  
*0 TO ±15 VOLTS*
- b) Pulse width:  
*4 TO 100 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*MIN PW: 0 TO 1.0 MHz*  
*MAX PW: 0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 VOLTS*  
*250 mA (MAX)*