


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVMM-1-P-C-OT-R*

S.N.: *3872*


Date: *July 16<sup>th</sup> 1967*



① *40 db ATTEN  
5 VOLTS/DIV  
OUT, 12 MHz*




② *AS ① BUT TR OUT  
POSITION 1*



③ *AS ② BUT TR OUT  
POSITION 5.*

- a) Output signal amplitude:  
*0 TO +2V  
(USEFUL TO +5V)*
- b) Pulse width:  
*1.0 TO 10 NSEC.*
- c) Rise time:  
*≤ 300 PSEC TO 2.0  
NSEC*
- d) Fall time:  
*≤ 600 PSEC TO 2.0 NSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V  
50-60 Hz*



④ *OUT, 25 MHz.  
40 db.*