


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ARR-A-1-PW-CA-N*

S.N.: *3883*

Date: *Jul 5 87*



- a) Output signal amplitude:  
*0 TO -200V.*
- b) Pulse width:  
*50 NSEC TO 1.0 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*MIN PW: 0 TO 100 KHZ*  
*MAX PW: 0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*R<sub>L</sub> = 50Ω*  
*5 KHZ.*

*+24 ± 1.0 VOLTS*  
*500mA (MAX)*

*[Signature]*