


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2-N*  
S.N.: *3893*  
Date: *AUG 7 87*



- a) Output signal amplitude:  
*0 TO -15V.*
- b) Pulse width:  
*0.2 TO 2.0 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN*  
*10 VOLTS/DIV*  
*≈ 12 MHz*

*+24 ± 1.0 VOLTS*  
*500 mA MAX*