


PULSE GENERATOR  
PERFORMANCE CHECK


Model: MPR-A-1-C-DN-CA-0W-0S

S.N.: 3896

Date: APR 20 87



Post  $R_L = 50 \Omega$   
5 KHz



Post  $R_L = 50 \Omega$   
5 KHz

- a) Output signal amplitude:  
0 TO  $\pm 200$  V
- b) Pulse width: TO 50 ns  
50 NSER TO 1.0 USER
- c) Rise time:  
 $\leq 10$  NSER
- d) Fall time:  
 $\leq 10$  NSER
- e) PRF:  
MIN PW: 0 TO 100 KHz  
MAX PW: 0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V  
50-60 Hz