

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


AVG-1-P

S.N.:

3898

Date:

SEPT 3 1987



~ 60 AMPERS/DIV  
500 Hz.

- a) Output signal amplitude:  
+25 TO +100 AMPERS
- b) Pulse width:  
20 ± 5 nsec
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

+15 ± 1.0 VOLT  
250 mA (max)