

PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVN-1-CPN-0T*

S.N.: *3900*

Date: *APR 20 87*



*(A) 40 db ATTEN
5.0 VOLTS DIV
≈ 100 MHz*



*(B) AS @ BUT
≈ 200 MHz*

- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*