


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3A-C-P*

S.N.: *3910*

Date: *OCT 8 87*



$R_L = 5.1 \Omega$

60 db ATTEN
20 AMPS/DIV
20 KHz

- a) Output signal amplitude:
+25 TO +75 AMPS
- b) Pulse width:
2 ± 0.5 NSEC
- c) Rise time: *(HALF POWER)*
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz