


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVM-2-P .

S.N.: 3937

Date: SEPT 17 87



- a) Output signal amplitude:  
0 TO +15 VOLTS
- b) Pulse width:  
0.2 TO 2.0 NSEC
- c) Rise time:  
 $\leq 100$  PSEC
- d) Fall time:  
 $\leq 135$  PSEC
- e) PRF:  
0 TO 25 MHZ
- f) Jitter, stability:  
OK
- g) Prime power:

+24  $\pm$  1.0 VOLTS

600 mA. (MAX)

40 db ATTEN  
 $\therefore$  10 VOLTS/DIV  
 $\approx$  5.0 MHZ