


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVM-2-P .

S.N.: 3940

Date: SEPT 17 87



- a) Output signal amplitude:
0 TO +15 VOLTS.
- b) Pulse width:
0.2 TO 2.0 NSEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 25 MHZ
- f) Jitter, stability:
OK
- g) Prime power:

+24 \pm 1.0 VOLTS

600 mA (MAX)

40 db ATTEN
 \therefore 10 VOLTS/DIV
 \approx 5.0 MHZ