


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PI-V-HV3-C-P*

S.N.: *3941*


Date: *OCT 14 '87*



Ⓐ *60 db ATTEN
100 VOLTS/DIV*



Ⓑ AS Ⓐ BUT 1 nSEC/DIV
(RISE TIME)



Ⓒ *20 VOLTS/DIV
d 200 PSEC/DIV (RISE TIME)*

- a) Output signal amplitude:
0 TO +100 V.
- b) Pulse width:
5 TO 100 nSEC.
- c) Rise time:
≤ 200 PSEC.
- d) Fall time:
≤ 4 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120 / 240 V
50-60 Hz*

[Signature]